



Declaration of Conformity
in accordance with EN ISO 17050-1:2010

We **Durham Geo Slope Indicator**
of 12123 Harbour Reach Drive Ste 106
Mukilteo WA 98275 USA

hereby declare that:

Equipment **Vibrating Wire Piezometers**
Model number **52611020-52611050, 52621020-52621050, 52610520-52610550,
52611610-52611625, 52612402, 52612405**

the product to which this declaration relates, is in conformity with the essential protection requirements of 2004/108/EC Electromagnetic Compatibility directive, as amended by EN61326-1, ed3. The products are in conformity with the following standards and/or other normative documents:

EMC:

Harmonized Standards: EN 61326-1:2006 Lab Equipment, EMC

IEC61000-6-3:2007 Emission standard for residential, commercial and light-industrial environments

IEC61000-4-2:2008 Electrostatic discharge immunity test

IEC61000-4-3:2006 Radiated, radio-frequency, electromagnetic field immunity test

IEC61000-4-4:2012 Electrical fast transient/burst immunity test

IEC61000-4-5:2005 Surge immunity test

IEC61000-4-6: 2008 Immunity to conducted disturbances, induced by radio-frequency fields

IEC61000-4-11:2004 Voltage dips, short interruptions and voltage variations immunity tests

I hereby declare that the equipment named above has been designed to comply with the relevant sections of the above referenced specifications. The items comply with all applicable Essential Requirements of the Directives

Signed by: *N Thatcher Harvey*

Name: N. Thatcher Harvey
Position: Engineering Manager
Done at: Mukilteo WA 98275 USA
On: 01 May 2014

Document ref. No.
201401011531